

PD233: Design of Biomedical Devices and Systems

(Lecture-5 Design Documentation)

Dr. Manish Arora

CPDM, IISc

Course Website:

<http://cpdm.iisc.ac.in/utsaah/courses/>

FDA CFR 21

In United States the requirements of Documentation are mandated by Title 21 of *Code of Federal Regulations (CFR)* for both drugs and medical devices.

- Volume 1: Parts 1–99 (FDA, General)
- Volume 2: Parts 100–169 (FDA, Food for Human Consumption)
- Volume 3: Parts 170–199 (FDA, Food for Human Consumption)
- Volume 4: Parts 200–299 (FDA, Drugs: General)
- Volume 5: Parts 300–499 (FDA, Drugs for Human Use)
- Volume 6: Parts 500–599 (FDA, Animal Drugs, Feeds and Related Products)
- Volume 7: Parts 600–799 (FDA, Biologics; Cosmetics)
- **Volume 8: Parts 800–1299 (FDA, Medical Devices)**
- Volume 9: Parts 1300–End (DEA and Office of National Drug Control Policy)

for example

21 CFR part 820: defines medical device quality system (QS) regulation;

-Part M: of the act defines general and specific record-keeping requirements for medical devices.

Documentation requirements

- Maintained by manufacturer
- Auditable by FDA or equivalent authority
- They must be legible and stored so as to minimize deterioration and to prevent loss
- Those stored digitally must be backed up and have a disaster plan in effect
- Can be deemed as confidential
- To be maintained for expected life time of the product (but not less than 2 year)

Documents

Information used to support an effective and efficient organizational operation.

Documents are created as a part of your planning

Need to be reviewed, updated and approved – i.e. document control.

- Business proposal
- Product specification
- Design specification
- Software quality assurance plan (SQAP) (where applicable)
- Software requirements specification (SRS) (where applicable)
- Software design description (SDD) (where applicable)

Records

Records are evidence of activities in past.

Records are facts and should not change, new records can be added.

Four types of records in context of medical devices:

1. Design history file (DHF)
2. Device master record (DMR)
3. Device history record (DHR)
4. Technical documentation file (TDF)

Documents

Business Proposal

Purpose: to identify and document market needs, market potential, the proposed product and product alternatives, risks and unknowns, and potential financial benefits.

The business proposal usually contains:

- Project overview, objectives, major milestones, schedule
- Market need and market potential
- Product proposal
- Strategic fit
- Risk analysis and research plan
- Economic analysis
- Recommendation to form a core project team
- Supporting documentation

Product Specification

- First step in realizing a product idea into a approved product development effort.
- It specifies ***what the product will do, how it will do it, and how reliable it will be***. To be effective, it must be as precise as possible.
- The product specification ***should be a controlled document***, i.e., subject to revision level control, so that any changes that arise are subjected to review and approval before implementation.

Product Specifications

The specification is a detailed review of the proposed product and includes

- Type of product
- Market it addresses
- Technology to be used
- Function of the product
- Product parameters necessary to function effectively
- Accuracy requirements
- Tolerances necessary for function
- Anticipated environment for the device
- Cautions for anticipated misuse
- Safety issues
- Human factors issues
- Anticipated life of the product
- Reliability goal
- Requirements from applicable domestic or international standards

Product Specifications

- ***Product specification are derived by agreement of multiple stakeholders:*** marketing, design engineering, manufacturing, customer service, reliability assurance, quality assurance, and regulatory affairs
- ***Each requirement should be identified with some form of notation,*** such as brackets and a number.

Example:

5.3.1 Analog to Digital Converter

The output of the analog to digital converter must be between X and Y [1].

This requirement can be later referred to as 5.3.1-1

Design Specifications

- The design specification is a document, which is derived from the product specification.

The requirements found in the product specification are partitioned and distilled down into specific design requirements ***for each subassembly.***

Design Specification

The design specification should address the following areas for each subsystem:

- Reliability budget
- Service strategy
- Manufacturing strategy
- Hazard consideration
- Environmental constraints
- Safety
- Cost budgets
- Standards requirements
- Size and packaging
- Power budget
- Heat generation budget
- Industrial design/human factors
- Controls/adjustments
- Material compatibility

Design Specifications

All electrical and mechanical ***inputs and outputs*** and their ***corresponding limits*** under all operating modes must be defined.

As in the product specification, the requirements in the ***design specification should be identified by a notation*** such as a bracket and numbers.

Other Controlled Documents

- Software quality assurance plan (SQAP)
- Software requirements specification (SRS)
- Software design description (SDD)

Reference: Chapter 5, King & Fries, Design of Biomedical Devices and Systems

Assignment 2

Prepare 5-7 minute presentation for business case your proposed team project device.

- Starting with **market need** and **market potential** your task is convince senior management of your organization that it is a product line getting into and your team is fit for doing the job (strategic fit).
- All data sources should be referenced
- **Major milestones, schedule** for you product/prototype development
- **Risk analysis** for the project should also be presented

Presentations on 6th September 2016 (10am)

Project Allocations

Records for Medical Device

1. Design history file (DHF)
2. Device master record (DMR)
3. Device history record (DHR)
4. Technical documentation file (TDF)

Design history file (DHF)

- DHF is a compilation of records, which describes the ***design history*** of a finished device.
- It covers the design activities used to develop the device, accessories, major components, labeling, packaging, and production processes.
- The DHF contains or references the records necessary to demonstrate that the design was developed in accordance with the approved design plans and the requirements of the quality system regulation.
- CFR 21 820.30(j) requires that each manufacturer establish and maintain a DHF for each type of device.

Design history file (DHF)

The ***design plans and subsequent updates*** should be part of the DHF.


In addition, the Quality System (QS) regulation specifically requires that

- ***Results of a design review***, including identification of the design, the date, and the individual(s) performing the review, shall be documented in the DHF.
- ***Design verification shall confirm that the design output meets the design input requirements.*** The results of the design verification, including identification of the design, method(s), the date, and the individual(s) performing the verification, shall be documented in the DHF.

Design history file (DHF)

Typical documents that may be in, or referenced in, a DHF include

- Design plans
- Design review meeting information
- Sketches
- Drawings
- Procedures
- Photos
- Engineering notebooks
- Component qualification information
- Biocompatibility (verification) protocols and data
- Design review notes
- Verification protocols and data for evaluating prototypes
- Validation protocols and data for initial finished devices
- Contractor/consultants information
- Product Documentation
- Parts of design output/DMR documents that show plans were followed
- Parts of design output/DMR documents that show specifications were met



Medical device design process as specified US- FDA requirements (source <http://fda.gov>)

Design history file (DHF)

- Manufacture of the device are beneficiary of DHF
- When problems occur during redesign and for new designs, the DHF has the institutional memory of previous design activities
- The DHF contains valuable verification and validation protocols that are not in DMR.

Device master record (DMR)

DMR is a compilation of those records containing the ***specifications and procedures for a finished device***.

DMR should include or refer to location of information about:

- ***Device specifications*** including appropriate drawings, composition, formulation, component, specifications, and software specifications
- ***Production process specifications*** including the appropriate equipment specifications, production methods, production procedures, and production environment specifications
- ***Quality assurance procedures and specifications*** including acceptance criteria and the quality assurance equipment used
- ***Packaging and labeling*** specifications, including methods and processes used
- ***Installation, maintenance, and servicing*** procedures and methods

Device history record (DHR)

DHR is the ***actual production records*** for a particular device.

It show the ***processes, tests, rework, etc.*** that the device went through.

The DHR should include or refer to the location of the following information:

- Dates of manufacture
- Quantity manufactured
- Quantity released for distribution
- Acceptance records which demonstrate the device is manufactured in accordance with the DMR
- Primary identification label and labeling used for each production unit
- Any device identification and control numbers used

Technical documentation file(TDF)

TDF contains all the relevant design data by means of which the product can be ***demonstrated to satisfy the essential safety requirements.***

It must include:

- General description of the product, including any planned variants.
- Design drawings, methods of manufacture envisaged and diagrams of components, subassemblies, circuits, etc.
- Descriptions and explanations necessary to understand the above mentioned drawings and diagrams and the operations of the product.
- ***Results of the risk analysis*** and a list of applicable standards applied.
- For products placed on the market in a sterile condition, a description of the method is used.
- ***Results of the design calculations and of the inspections carried out.***
- ***Test reports and, where appropriate, clinical data.***
- Labels and instructions for use.

Quiz

	DHF	DMR	DHR	TDF
Clinical Trial Information				
Project Plans				
Design Specification				
Design Validation Plans				
Product Specification				
Sales order report				
Shipping report				
Risk analysis				
Tooling specs				
Installation Instructions				